

DECLARATION OF CONFORMITY**We, Haag-Streit UK,**

Unit C, Woodside Industrial Estate,
Dunmow Road, Bishop's Stortford,
Hertfordshire CM23 5RG,
United Kingdom
Tel: +44 (0) 1279 414969
Email: Regulatoryhsuk@haag-streit-uk.com

Declare under our sole responsibility that the product(s) listed below to which this declaration relates, is in conformity with the applicable provisions as required by Part II of UK Medical Devices Regulations 2002 (as amended) for general medical devices (as modified by Part II of Schedule 2A to the UK MDR 2002).

- Device Group:** Tonometer Calibration Weights
- Description:** Refer to Product Schedule on page 2 onwards for a complete list of part codes and descriptions
- Intended Purpose:** A reusable medical device accessory intended to be used to provide a known force to the reading scale on the Perkins Mk-3 Tonometer. Contact is transient use with the user's hands.
- Risk Classification:** The device(s) has been classified as **class I, Rule 1** in accordance with Part II of the UK MDR 2002, Annex IX (as modified by Part II of Schedule 2A to the UK MDR 2002).
- Designated Standards:** Refer to HSHS2017913 for a complete list of applied standards

Authorised By:



Signed:

guerino ostacchini (Jun 22, 2023 09:13 EDT)


Date:

22 Jun 2023

Mr G Ostacchini

Managing Director (Operations)

Signed:


Kerrie Mouncey (Jun 22, 2023 14:09 GMT+1)

Date:

22 Jun 2023

Kerrie Mouncey

QA/RA Manager

	Doc N° DCHS20221940	V3	Page 2 of 2
	Title; UK MDR Declaration of Conformity – Tonometer Calibration Weights		

Description	Part Code	GMDN	GTIN
5g & 2g Pair of Spare Weights	5801004-5	17407	05055296024846